

Life+ FutMon - Working Group on QA/QC in Laboratories
Meeting of the Heads of the Laboratories
12-13 October 2009 in Warsaw

***Analytical aspects comparison
DTN and IC determinations***

Gabriele Tartari¹ & Carmen Jacoban²

¹C.N.R. Institute of Ecosystem Study, Verbania (Italy)

²Forest Research Station, Campulung Moldovenesc (Romania)



FutMon



DTN determination

Tot-N UV220 pag. 1

Consiglio Nazionale delle Ricerche
INSTITUTE FOR ECOSYSTEM STUDY
Verbania Pallanza - Italy

Hydrochemistry Laboratory, Analytical Methods - Internal Use.
Gabriele TARTARI



Total Nitrogen in Water

A peroxodisulphate oxidation procedure followed by spectrophotometric determination

PRINCIPLE

Ammonium, nitrite and organic nitrogen, are oxidised to nitrate using potassium peroxodisulphate in a buffer boric acid-sodium hydroxide buffer. The oxidation of the nitrogen compounds is performed in an autoclave at 120 °C, resulting in a pH change of the buffer from 9.7 to 5.0. The resulting nitrate is determined by spectrophotometry at 220 nm. (Note: Organic substances which interfere at this wavelength are assumed to be completely oxidised in the peroxodisulphate digestion procedure).




Fig. 1 - Absorption spectrum of a nitrate solution between 200 nm to 250 nm

FutMOn



DTN determination

Total nitrogen UV determination 220 nm OP 1 cm from 2000 to 2009 CNR-ISE Italy

	Standard 1	Standard 2	Standard 3	Standard 4	Standard 5	Standard 6	Standard 7	Standard 8	Standard 9
NT 0.4-6 Media	0.094	0.188	0.237	0.353	0.470	0.700	0.934	1.154	1.381
Med.+2SD	0.101	0.196	0.251	0.372	0.491	0.725	0.978	1.211	1.430
Med.-2SD	0.088	0.179	0.222	0.334	0.448	0.675	0.889	1.098	1.332
R.S.D	3.6	2.3	3.1	2.7	2.3	1.8	2.4	2.4	1.8
n° dati	25	23	22	20	24	23	24	20	24
mg N/L	0.4	0.8	1.0	1.5	2.0	3.0	4.0	5.0	6.0
RF medio	0.2361	0.2349	0.2367	0.2350	0.2348	0.2333	0.2334	0.2308	0.2302

Total nitrogen UV determination 220 nm OP 1 cm - Carmen Jacoban 4-5 May 2009


	Standard 1	Standard 2	Standard 3	Standard 4	Standard 5	Standard 6	Standard 7	Standard 8	Standard 9
NT 0.4-6 Media	0.111	0.201	0.240	0.358	0.476	0.697	0.926	1.154	1.406
mg N/L	0.4	0.8	1.0	1.5	2.0	3.0	4.0	5.0	6.0
RF medio	0.2786	0.2509	0.2405	0.2384	0.2381	0.2323	0.2314	0.2308	0.2343

Diff.	-0.017	-0.013	-0.004	-0.005	-0.006	0.003	0.008	0.000	-0.025
--------------	---------------	---------------	---------------	---------------	---------------	--------------	--------------	--------------	---------------




FutMon

DTN calibrations comparison



CNR-ISE Italy
Average 10 years
calibrations 2000-2009



Carmen Jacoban
May 2009
One calibration



FutMon

IC determination

Instruments:

Dionex ICS 3000 cations

Dionex DX 500 cations

Dionex DX 320 anions

Dionex ICS 3000




Gabriele Tartari

Carmen Jacoban




FutMon


Cloruri



Azoto nitrico (Detector Cond. & UV)



Solfati




CNR-ISE control chart from Jul 2008-Apr 2009

○ Carmen's data 2 May 2009




FutMon


Sodio




Magnesio




Azoto ammoniacale



Potassio



Calcio



— UCL - - - UWL - - - LWL — LCL —■— Dati —— Media

UCL

Upper Control Limit (massimo accettato)

= Media + 3 D.S. consigliata

UWL

Upper Warning Limit (massimo consigliato)

= Media + 2 D.S. consigliata

LWL

Lower Control Limit (minimo consigliato)

= Media - 2 D.S. consigliata

LCL

Lower Warning Limit (minimo accettato)

= Media - 3 D.S. consigliata

CNR-ISE control chart from Mar-Apr 2009



Carmen's data 2 May 2009




FutMon

Thank you for your attention




FutMon





CNR-ISE Italy



N-NH₄ comparison SPEC Phe - IC

Jen 2005 - Jun 2006

Total Range (mg L ⁻¹)	0.05 - 2.00	
Half Range on mediane (mg L ⁻¹)	0.050 - 0.30	0.30 - 2.00
N. data	476	508
Test Wilcoxon (a 0.001)		
Probability	OK 0.037	OK 0.78
IC = a + b x SPEC		
r	0.9316	0.9924
Test Slope = 1 (0.001)		
Slope	OK 0.989	NON OK 1.021
Test Intercept = 0 (0.001)		
Intercept	OK 0.003	OK -0.014
DIFFERENCE from data	OK	OK
DIFFERENCE (µg N L⁻¹)	SPEC - IC	
Min	-98	-176
25 %ile	-16	-29
Mediane	-3	-1
75 %ile	12	29
Max	97	128
Average	-2	0
DIFFERECE absolute (µg N L⁻¹)		
25 %ile	7	13
Mediane	15	29
75 %ile	26	52
Max	98	176
Average	20	34

CNR-ISE Italy

